


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130292</u>	
Name	<u>Andie Arhee</u>	ID#	<u>2385</u> Date <u>7-29-09</u>
A. Agency	<u>Bethel Police Dept</u>		Phone # <u>907-543-381</u>
Instrument Location	<u>BPD SQUAD ROOM</u>		
Alco S/N <u>X124650</u>		Target Value <u>.086</u>	High Pressure <u>1100</u>
B. Alco Test Values			
<u>.084</u>		<u>.083</u>	
1 st Alco		2 nd Alco	
Signature			
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14 day of October, 2009.

 (Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office 11/15/2011

Lori A Zirkle

BMB
10/14/09

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130292

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130292

JULY 29, 2009

OPERATOR'S NAME:

ACHEE/ANDRE

OPERATOR'S NUMBER: 2385

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: XXXXXX

DEPT/AGENCY: BETI

CASE/REPORT: XXXXXX

TEST TYPE: T

ALCO TARGET VALUE: .006

ALCO S/N: X124650

--- BREATH ANALYSIS ---

.006 ADJUSTED FOR 29.9% in
ALCO TARGET .006 17:44
BLANK TEST .000 17:45
INTERNAL STANDARD VERIFIED 17:45
ALCO TU 29.9% in .004 17:45
BLANK TEST .000 17:46
SUBJECT SAMPLE .000 17:46
BLANK TEST .000 17:47
ALCO TU 29.9% in .003 17:47
BLANK TEST .000 17:48

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130292

JULY 29, 2009

TIME 17:41

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.96 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~